

Surgical Face Mask



- CE
- CFDA
- COVID-19 Medical Device

Authorization for
Importation or Sale in
Canada

- White List

Material Three-layer Mask
 Inner layer: Hygroscopic nonwoven fabric
 Protective layer: 25gms Melt-blown fabric
 Outer later: Waterproof nonwoven fabric

Size 17.5*9.5cm

Standard EN 14683:2019

Model Name	Type I (Ear loop)	Type II (Ear loop)	Type IIR (Ear loop)	Type IIR (Tie on)
Cat. No.	O06B010110	O06B020110	O06B020210	O06B020310
Bacterial filtration efficiency	>95%	>98%	>98%	>98%
Differential pressure	<40Pa/cm ²	<40Pa/cm ²	<60Pa/cm ²	<60Pa/cm ²
Splash resistance pressure	/	/	≥16kPa (120mmHg)	≥16kPa (120mmHg)

WELL LEAD MEDICAL CO., LTD.

Address: 47 Guomao Avenue South, Panyu, Guangzhou, China
 Tel: +86-20-84758878 Fax: +86-20-84758224 Post Code: 511434
 Marketing & Trading Office
 Add: RM 805, No 55, 21 Long, Pusan Road, Shanghai, China
 Tel: +86-21-58397623 Fax: +86-21-58397615 Post code: 200125
 E-mail: info@welllead.com.cn http://www.welllead.com.cn

wellead[®]